



---

## Journal Articles

---

2020

# Inpatient Use of Ambulatory Telemetry Monitors for COVID-19 Patients Treated With Hydroxychloroquine and/or Azithromycin

D. Chang

*Northwell Health*, [dchang4@northwell.edu](mailto:dchang4@northwell.edu)

M. Saleh

*Northwell Health*, [msaleh2@northwell.edu](mailto:msaleh2@northwell.edu)

J. Gabriels

*Northwell Health*, [jgabriel13@northwell.edu](mailto:jgabriel13@northwell.edu)

H. Ismail

*Zucker School of Medicine at Hofstra/Northwell*, [hismail@northwell.edu](mailto:hismail@northwell.edu)

B. Goldner

*Zucker School of Medicine at Hofstra/Northwell*, [bgoldner@northwell.edu](mailto:bgoldner@northwell.edu)

*See next page for additional authors*

Follow this and additional works at: <https://academicworks.medicine.hofstra.edu/articles>



Part of the [Cardiology Commons](#)

---

## Recommended Citation

Chang D, Saleh M, Gabriels J, Ismail H, Goldner B, Willner J, Beldner S, Mitra R, John R, Epstein LM. Inpatient Use of Ambulatory Telemetry Monitors for COVID-19 Patients Treated With Hydroxychloroquine and/or Azithromycin. . 2020 Jan 01; 75(23):Article 6048 [ p.]. Available from: <https://academicworks.medicine.hofstra.edu/articles/6048>. Free full text article.

This Article is brought to you for free and open access by Donald and Barbara Zucker School of Medicine Academic Works. It has been accepted for inclusion in Journal Articles by an authorized administrator of Donald and Barbara Zucker School of Medicine Academic Works. For more information, please contact [academicworks@hofstra.edu](mailto:academicworks@hofstra.edu).

---

**Authors**

D. Chang, M. Saleh, J. Gabriels, H. Ismail, B. Goldner, J. Willner, S. Beldner, R. Mitra, R. John, and L. M. Epstein



Since January 2020 Elsevier has created a COVID-19 resource centre with free information in English and Mandarin on the novel coronavirus COVID-19. The COVID-19 resource centre is hosted on Elsevier Connect, the company's public news and information website.

Elsevier hereby grants permission to make all its COVID-19-related research that is available on the COVID-19 resource centre - including this research content - immediately available in PubMed Central and other publicly funded repositories, such as the WHO COVID database with rights for unrestricted research re-use and analyses in any form or by any means with acknowledgement of the original source. These permissions are granted for free by Elsevier for as long as the COVID-19 resource centre remains active.

Inpatient Use of Ambulatory Telemetry Monitors for COVID-19 Patients Treated with Hydroxychloroquine and/or Azithromycin

David Chang, MD, Moussa Saleh, MD, James Gabriels, MD, Haisam Ismail, MD, Bruce Goldner, MD, Jonathan Willner, MD, Stuart Beldner, MD, Raman Mitra, MD, PHD, Roy John, MBBS, PHD, Laurence M. Epstein, MD

PII: S0735-1097(20)35009-9

DOI: <https://doi.org/10.1016/j.jacc.2020.04.032>

Reference: JAC 27285

To appear in: *Journal of the American College of Cardiology*

Received Date: 2 April 2020

Revised Date: 12 April 2020

Accepted Date: 16 April 2020

Please cite this article as: Chang D, Saleh M, Gabriels J, Ismail H, Goldner B, Willner J, Beldner S, Mitra R, John R, Epstein LM, Inpatient Use of Ambulatory Telemetry Monitors for COVID-19 Patients Treated with Hydroxychloroquine and/or Azithromycin, *Journal of the American College of Cardiology* (2020), doi: <https://doi.org/10.1016/j.jacc.2020.04.032>.

This is a PDF file of an article that has undergone enhancements after acceptance, such as the addition of a cover page and metadata, and formatting for readability, but it is not yet the definitive version of record. This version will undergo additional copyediting, typesetting and review before it is published in its final form, but we are providing this version to give early visibility of the article. Please note that, during the production process, errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

© 2020 Published by Elsevier on behalf of the American College of Cardiology Foundation.

## **Inpatient Use of Ambulatory Telemetry Monitors for COVID-19 Patients Treated with Hydroxychloroquine and/or Azithromycin**

David Chang, MD<sup>1</sup>, Moussa Saleh, MD<sup>1</sup>, James Gabriels, MD<sup>1</sup>, Haisam Ismail, MD<sup>1</sup>, Bruce Goldner<sup>1</sup>, MD, Jonathan Willner, MD<sup>1</sup>, Stuart Beldner, MD<sup>1</sup>, Raman Mitra, MD, PHD<sup>1</sup>, Roy John, MBBS, PHD<sup>1</sup>, Laurence M. Epstein, MD<sup>1</sup>

1. Northwell Health, North Shore University Hospital, Department of Cardiology, Division of Electrophysiology, Manhasset, New York, USA

**Short Title:** Inpatient Use of Ambulatory Monitors for COVID-19 Patients

**Conflict of Interest:** None

### **Corresponding Author:**

David Chang  
300 Community Drive  
Manhasset, New York 11030  
Phone: +1-516-306-7473  
Fax: +1-516-562-4882  
Email: [davidchang7787@gmail.com](mailto:davidchang7787@gmail.com)

**Keywords:** COVID-19, coronavirus, ambulatory monitor, hydroxychloroquine, azithromycin

AF=Atrial fibrillation; AZM=Azithromycin; COVID-19=Coronavirus 2019;  
ECG=Electrocardiogram; HCP=Healthcare provider; HCQ=Hydroxychloroquine;  
MCOT=Mobile cardiac outpatient telemetry; PPE=Personal protective equipment;  
QTc=Corrected QT

Coronavirus 2019 (COVID-19) has led to a rapid increase in hospital admissions, placing stress on healthcare systems that have a finite number of hospital beds, healthcare providers (HCPs), and medical supplies. Preliminary data suggest that hydroxychloroquine (HCQ) and azithromycin (AZM) may improve the clinical course in patients with COVID-19. (1,2) However, HCQ±AZM may increase the risk for arrhythmias and sudden cardiac death due to QT prolongation. (3,4) Given the widespread use of HCQ±AZM, it is challenging to monitor all of these inpatients on telemetry. Performing serial electrocardiograms (ECGs) for QTc monitoring increases HCP exposures and personal protective equipment (PPE) use.

We placed Mobile Cardiac Outpatient Telemetry (MCOT) (BioTelemetry, Malvern, PA, USA) on patients receiving HCQ±AZM for COVID-19 on non-telemetry floors. Following a baseline ECG, subsequent ECGs were cancelled. Telemetry technicians applied the MCOTs and linked them to the device phone. Patients had bidaily QTc measurements while receiving HCQ±AZM. An electrophysiologist received “urgent alerts” and bidaily reports from BioTelemetry. A QTc>500ms and any arrhythmias generated “urgent alerts.” If a patient was discharged to complete HCQ±AZM as an outpatient or remained hospitalized after completing HCQ±AZM, the MCOT was removed, sterilized, and reused.

In one week, 117 consecutive COVID-19-positive patients on HCQ±AZM without telemetry monitors received an MCOT. The average age was 60.2±14.9 years (range 27-93 years); 40.5% were female, 52.1% had hypertension, 28.2% had diabetes, 0.9% had heart failure, and 5.1% had coronary artery disease. All patients were treated with HCQ 400mg bidaily for one day followed by 200mg bidaily for four days. Fifty-one (43.6%) patients also received ≥1 doses of intravenous azithromycin 500mg. Forty (34.2%) patients also received ≥1 other QT prolonging medications. Over the course of 295 total patient days, there were 28 urgent alerts

for 18 (15.4%) patients. Atrial fibrillation (AF) with a rapid ventricular response was the most common (15, 53.6%). There were five (17.9%) alerts for  $QTc > 500ms$  (**Table 1**). An electrophysiologist was contacted for urgent events within 3-5 minutes. Of the 28 urgent alerts, 12 did not warrant intervention (e.g. first-degree atrioventricular block).

From a baseline mean  $QTc$  of  $437.1 \pm 22.2ms$ , the average increase in  $QTc$  for the entire population was  $33.9 \pm 26.8ms$  (Table 1). The maximum  $QTc$  was similar in patients treated with HCQ versus HCQ+AZM ( $448.5 \pm 33.7ms$  vs.  $451.9 \pm 29.2ms$ ,  $p=0.58$ ). The change in  $QTc$  from baseline was also similar ( $32.1 \pm 25.1ms$  vs.  $35.7 \pm 28.9ms$ ,  $p=0.66$ ). HCQ was discontinued in one patient after three days due to  $QTc$  prolongation from 460ms to 565ms.

This study demonstrates that when hospital admission rates exceed the capacity of telemetry beds, the MCOT may be used to monitor for arrhythmias and assess the  $QTc$ . In 2017, the MCOT, which consists of a sensor and monitor network that communicate via Bluetooth, was FDA-approved for  $QTc$  measurement, analysis, and reporting. Once gathered, the data is forwarded to the monitor for analysis. After each use, MCOT may be rapidly “redeployed” to another patient.

In our experience, twenty-eight “urgent alerts” were communicated in near real-time to an electrophysiologist, of which 16 alerts resulted in management changes. In addition to the “urgent alerts,” the MCOT afforded electrophysiologists the ability to monitor for  $QTc$  changes. While HCQ±AZM may put patients at higher risk for drug-induced arrhythmias, none of our patients had arrhythmias that led to medication discontinuation. The MCOT also allowed for better utilization of HCPs and resources. By eliminating the need for serial ECGs, we reduced both HCP exposures and PPE use.

The limitations of MCOT include that the device was never approved to measure QTc for patients with AF or atrial flutter, QRS >160ms, and T-wave <5% of the peak QRS amplitude. The single-center, non-randomized study design, and a healthy population from a cardiac standpoint, are other limitations. The MCOT must be used with caution in patients with significant cardiac disease.

In conclusion, innovative management of COVID-19 patients treated with HCQ±AZM is needed given the limited healthcare resources. The MCOT may be utilized for arrhythmia and QTc monitoring while reducing both HCP exposures and PPE use.



**References:**

1. Gautret P, Lagier JC, Parola P, et al. Hydroxychloroquine and azithromycin as a treatment of COVID-19: results of an open-label non-randomized clinical trial. *Int J Antimicrob Agents* – In Press DOI:10.1016/j.ijantimicag.2020.105949.
2. Chen Z, Hu J, Zhang Z et al. Efficacy of hydroxychloroquine in patients with COVID-19: results of a randomized clinical trial. *medRxiv* 2020.03.22.20040758; doi:<https://doi.org/10.1101/2020.03.22.20040758>.
3. Chen CY, Wang FL, Lin CC. Chronic hydroxychloroquine use associated with QT prolongation and refractory ventricular arrhythmia. *Clin Toxicol (Phila)* 2006;44:173–175.
4. Ray WA, Murray KT, Hall K, Arbogast P, Stein CM, Azithromycin and the risk of cardiovascular death. *New Engl J Med* 2012;366:1881–1890.

**Table 1: MCOT Urgent Alerts and QTc Measurement of the Study Cohort**

Urgent Alerts (N=28)				
Atrial Fibrillation with a Rapid Ventricular Response			15 (53.6%)	
QTc >500ms			5 (17.9%)	
First Degree Atrioventricular Block			4 (14.3%)	
Nonsustained Ventricular Tachycardia			2 (7.1%)	
Ventricular Bigeminy			1 (3.6%)	
Supraventricular Tachycardia			1 (3.6%)	
QTc Measurements				
QTc Parameters	Overall (N=117)	HCQ (N=66)	HCQ+AZM (N=51)	P-Value
Baseline QTc	437.1 ± 22.2	438.1 ± 23.8	435.8 ± 19.9	0.591
Maximal QTc	449.9 ± 31.7	448.5 ± 33.7	451.9 ± 29.2	0.575
Maximal QTc – Baseline QTc	33.9 ± 26.9	32.1 ± 25.1	35.7 ± 28.9	0.662

Final QTc	441.2 $\pm$ 28.7	440.0 $\pm$ 32.1	443.3 $\pm$ 23.6	0.54
Final QTc – Baseline QTc	7.3 $\pm$ 30.7	3.9 $\pm$ 31.9	12.8 $\pm$ 29.3	0.247

Values listed are numbers (percentages) or means $\pm$ standard deviations. AZM=Azithromycin; HCQ=Hydroxychloroquine; QTc=Corrected QT.